



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/541,297

06/30/2005

Jurgen Kress

6636

60333

7590

04/12/2010

EDWIN D. SCHINDLER  
FIVE HIRSCH AVENUE  
P.O. BOX 966  
CORAM, NY 11727-0966

EXAMINER

NIA, ALIREZA

ART UNIT

PAPER NUMBER

3739

MAIL DATE

DELIVERY MODE

04/12/2010

PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

<b>Office Action Summary</b>	<b>Application No.</b> 10/541,297	<b>Applicant(s)</b> KRESS, JURGEN	
	<b>Examiner</b> ALIREZA NIA	<b>Art Unit</b> 3739	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

### Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 02 March 2010.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 29-48 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 29-48 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 30 June 2005 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All    b) ☐ Some \*    c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☒ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- |   |   |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)         | 4) <input type="checkbox"/> Interview Summary (PTO-413)           |
| 2) <input type="checkbox"/> Notice of Draftperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____                                      |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08)         | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____   | 6) <input type="checkbox"/> Other: _____                          |

## **DETAILED ACTION**

### ***Continued Examination Under 37 CFR 1.114***

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on March 2<sup>nd</sup>, 2010 has been entered.

### ***Claim Rejections - 35 USC § 102***

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

3. **Claims 29-38, 42, and 47-48 rejected under 35 U.S.C. 102(b) as being anticipated by Silverstein 4,646,722.**

4. Regarding claims 29-38, and 42, Silverstein discloses an endoscope 8 with hygiene protection 10,30, comprising a cover 10,30,48 closed at a distal end 32 and transmissible for optical information via 32,44,46, at least on a front face 32 of said cover 10,30,48, and capable of being rolled via 48 in an axial direction of said endoscope 8, a working channel 38,36 extending parallel to said endoscope 8 via 21 and terminating in an open mode via 38,40,42 at said distal end 32 of said cover 10,30, said working channel 36,38 being connected only to said distal end 32 of said cover 10,30,48 and positioned between an outer side 12 of said endoscope 8 via 21 and an inside of said cover 10,30,48 (col. 6, lines 16-46, 58-68, col. 7, lines 1-14, 41-54,

Art Unit: 3739

figs. 1-3), and consisting of only a single vacuum channel 64 (figs. 5-6, col. 7, lines 60-65), having at least one opening A, and terminating at said inside of said cover 10,30,48 with said at least one opening A terminating inside of said cover 10,30,48 in a direction facing said endoscope 8 (the direction is facing endoscope 8 and portion 14 of endoscope 8), said single vacuum channel 64 being a dedicated channel 64 that is a different channel 64 from said working channel 36,38 (fig. 4). Silverstein further discloses means 48 (the elastomeric material) for varying cross-sectional diameter via 48,60,64 (col. 7, lines 47-60), wherein said cover 10,30,48 includes a material 48 that is airtight, watertight and impermeable to pathological microorganisms with connection of said cover 10,30,48 to said working channel 36,38 to said distal end 32 of said cover 10,30,48 being made airtight, watertight and impermeable to pathological microorganisms (col. 6, lines 35-46), wherein said cover 10,30,48 is flexible and elastic and foldable in the axial direction of said endoscope 12 via 48 (col. 6, lines 58-68), wherein at least a portion of said cover 10,30,48 inherently has an internal diameter that is larger than an external diameter 12 of the endoscope 8. Silverstein further discloses a transparent pane or lens 32,44,46 on said distal end 32 of said cover 10,30,48 on said front face 32 of said cover 10,30 (col. 6, lines 47-57), wherein said transparent pane or said lens 32,44 at least partially forms said front of face 32 of said distal end 32 of said cover 10,30,48 via 46, wherein said distal end 32 of said cover is an optically transparent cap via 32,44, wherein said distal end 30,32 of said cover 10,30,48 has a wall thickness that is greater than wall thickness of a non-distal region 48 of said cover 10,30,48, wherein said cover 10,30,48, when open at a proximal end 48, is fixable to be airtight on a shaft 12,14 of said endoscope 8 via 48,66 (col. 6, lines 62-68).

Silverstein further discloses means (suction, fig. 4) for applying sub-atmospheric pressure to said

Art Unit: 3739

single vacuum channel 64 during use of said endoscope 8 (col. 6, lines 28-35). Silverstein further discloses depressions 21 in an axial direction on an outer surface of said endoscope 8 (col. 6, lines 16-21).

5. With respect to claims 47-48, Silverstein discloses attaching a hygiene protection 10,30 to an endoscope 8, said endoscope 8 including a cover 10,30,48 closed at a distal end 32 and transmissible for optical information via window 32,44, at least on a front face 32 of said cover 10,30,48, and capable of being rolled via 48 in an axial direction of said endoscope 8, a working channel 36,38,39 extending parallel to said endoscope 8 via 21 and terminating in an open mode 42 via 40 at a distal end 32 of said cover 10,30,48, said at least one working channel 36,38,39 being connected only to said distal end 32 of said cover 10,30,48 and positioned between an outer side 12 of said endoscope 8 via 21 and an inside of said cover 10,30,48, and consisting of only a single vacuum channel 64, having at least one opening A (see inserted image 1 below), and terminating via A at said inside of said cover 10,30,48 via 48 in a direction facing said endoscope 8 via 14 (the direction is facing endoscope 8 and portion 14 of endoscope 8), said single vacuum channel 64 being a dedicated vacuum channel 64 that is a different channel from said working channel 36,38 (col. 6, lines 30-46, fig. 3), said method comprising the steps of coating an inner side of said window 32,44 at said distal end 32 of said cover 10,30 for producing optical contact between said window 32,44 and an optical channel of said endoscope (col. 6, lines 52-57), introducing a distal end of said endoscope 8 into said cover 10,30,48, said cover being open at a proximal end 48 and closed at said distal end 32 thereof, rolling said cover 48 onto, or unfolding said cover 48 with, an enclosure 12 of said endoscope 8 and said working channel 36,38 via 21, and applying a sub-atmospheric pressure via (suction, see fig. 5-6, col. 7,

Art Unit: 3739

lines 60-65) to said single vacuum channel 64, fixing said working channel 36,38 at said distal end 32 of said cover 10,30,48 via 40,42, said working channel 36,38 being positioned in depressions 21 provided on an outer surface 12 of said endoscope 8 and said working channel 36,38 via 35.

**6. Claims 29, 41, and 46 rejected under 35 U.S.C. 102(b) as being anticipated by Sidall 4,741,326.**

7. Sidall discloses an endoscope (fig. 2) with hygiene protection1, comprising a cover 1 closed at a distal end 3 and transmissible for optical information (col. 2, lines 48-65), at least on a front face 3 of said cover 1, and capable of being rolled in an axial direction of said endoscope, a working channel 19 extending parallel to said endoscope (col. 3, lines 60-68) and terminating in an open mode at said distal end 3 of said cover 1, said working channel 19 being connected only to said distal end 3 of said cover 1 and positioned between an outer side of said endoscope (col. 3, lines 56-60) and an inside of said cover 1, and consisting of only a single vacuum channel 5, having at least one opening (at the distal and proximal ends thereof), and terminating at said inside of said cover (via it's proximal end) with said at least one opening terminating inside of said cover 1 in a direction facing said endoscope (i.e. in the direction of 21), said single vacuum channel 5 being a dedicated vacuum channel 5 that is a different channel 5 from said working channel 19 (col. 2, lines 67-68, col. 3, lines 1-3), wherein said single vacuum channel 5 extends for at least a portion of a length of said cover 1 (fig. 5), wherein said working channel 19 and said single vacuum channel 5 are fixed to said distal end 3 of said cover 1 (fig. 5).

Art Unit: 3739

***Claim Rejections - 35 USC § 103***

8. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

**9. Claims 29 and 43-45 are rejected under 35 U.S.C. 103(a) as being unpatentable over Silverstein 4,646,722 in view of Kuramoto 5,630,795.**

10. Silverstein discloses an endoscope 8 with hygiene protection 10,30,48, comprising a cover 10,30,48 closed at a distal end 32 and transmissible for optical information via 32,44, at least on a front face 32 of said cover 10,30,48, and capable of being rolled via 48 in an axial direction of said endoscope 8, a working channel 35 extending parallel to said endoscope 8 and terminating in an open mode via 36,42 at said distal end 32 of said cover 10,30,48, said working channel 35 being connected only to said distal end 32 of said cover 10,30,48 and positioned between an outer side 12 of said endoscope 8 and an inside of said cover 10,30,48 via 21 (col. 6, lines 16-46, col. 6, lines 58-68, col. 1-14, col. 7, lines 41-54, figs. 1-3). Silverstein further discloses depressions 21 in an axial direction on an outer surface of said endoscope 8, said depression corresponding in shape and in depth to a diameter and profile of said working channel 35,38 (col. 6, lines 16-21).

11. However, Silverstein fails to positively disclose a single vacuum channel, having at least one opening, and terminating at said inside of said cover with said at least one opening terminating inside of said cover in a direction facing said endoscope, said single vacuum channel being a single dedicated channel that is a different channel from said working channel.

Art Unit: 3739

Silverstein further fails to positively disclose said depressions correspond in shape and in depth to a diameter and profile of said vacuum channel, wherein said depressions have a width, running in the axial direction on said outer surface of said endoscope.

12. Kuramoto teaches a vacuum channel 122,344, having at least one opening 125 with said at least one opening 125 terminating in a direction facing said endoscope 101,113, said vacuum channel 122,344,361,362 being a single dedicated vacuum channel that is a different channel from any other working channel (col. 6, lines 14-35, col. 7, lines 20-67, col. 12, lines 29-53, 63-67, col. 13, lines 1-13, col. 25, lines 37-64, figs. 17,44a,44b45a,45b) in use with an analogous endoscope with hygiene protection. Kuramoto further teaches depressions 352,367 corresponding in shape and in depth to a diameter and profile of said vacuum channel 344,361,362, wherein said depressions 352,367 have a width, running in the axial direction on said outer surface of said endoscope 351,365 which is smaller than the width of said depressions 352,367 at their center points (figs. 44a,b,45a,45b), making it possible to not only clean the forward-end portion of the endoscope to ensure disinfection but also allowing to reduce the diameter of the inserting section of the endoscope (col. 25, lines 55-64).

13. It would have been obvious to one of ordinary skill in the art at the time of the invention to have modified the endoscope with hygiene protection of Silverstein with the above discussed limitations as taught by Kuramoto in order to have provided an improved endoscope with hygiene protection that has an endoscope having a narrow inserting section that makes it possible to comfortably clean the forward-end portion of the endoscope to ensure disinfection of the endoscope distal tip while inside of said cover by providing ample space for maneuverability between the endoscope inserting section and the inside of the hygiene protection cover.



Art Unit: 3739

14. With respect to the recitation “vacuum channel, having at least one opening, and terminating at said inside of said cover” in claim 29, it would have been obvious to one of ordinary skill in the art at the time of the invention to recognize that based on the above discussed modification of the Silverstein reference in view of Kuramoto that the vacuum channel 122,344,361 with at least one opening 125, would have terminated inside of said cover with said at least one opening 125 terminating inside of said cover 10,30,48 in a direction facing said endoscope 101,113 to clean the forward-end portion of the endoscope to ensure disinfection of the endoscope distal tip while inside of said cover, wherein said vacuum channel 122,344,361 would also have been detachable from said distal end of said cover 10,30 since 122,344,361 is removable.

15. Regarding the recitation “said working channel...detachably connected to said distal end of said cover” in claim 45, it would have also been obvious to one ordinary skill in the art at the time of the invention to have removed working channel 35 via 40,42 in order to facilitate simple replacement of damaged working channels, since separation of elements, where removability would be desirable, is a design consideration within the skill of the art.

**16. Claim 39 is rejected under 35 U.S.C. 103(a) as being unpatentable over Silverstein 4,646,722 in view of Adair 5,433,221.**

17. Silverstein discloses the invention as discussed above.

18. However, Silverstein fails to positively disclose the cover to be conically enlarged in a vicinity of said proximal end with a portion of said cover being folded backwardly to be wrinkle-free in said vicinity of said proximal end and fixable via a chemically inert and non-toxic adhesive.

Art Unit: 3739

19. Adair teaches an analogous cover 10 used in covering an endoscopic surgical camera, the cover 10 having conically enlarged in a vicinity of its proximal end 12 (the distal end 14 having a smaller diameter than proximal end 12, col. 4, lines 33-42). The cover also having a portion 20 that is folded backwardly (col. 4, lines 55-59) wherein the proximal end 12 is fixable via an adhesive 18 (col. 4, lines 50-53) resulting in an improved sterile cover for an unsterile endoscope enhancing operating room procedures.

20. It would have been obvious to one of ordinary skill in the art at the time of the invention to have provided the device of Silverstein with above discussed limitations as taught by Adair in order to have an improved sterile cover for an unsterile endoscope enhancing operating room procedures by minimizing sterility problems with an apparatus of this type that requires a minimal amount of handling to be put in use.

**21. Claim 40 is rejected under 35 U.S.C. 103(a) as being unpatentable over Silverstein 4,646,722 in view of Darras 4,886,049.**

22. Silverstein discloses the invention as discussed above.

23. However, Silverstein fails to disclose a tear thread connected to said cover at said distal end and running parallel to said endoscope on the inside of said cover.


24. Darras discloses a tear thread 20 attached to an endoscope cover 10 (col. 3, lines 1-20) and running parallel to said endoscope on an inside of the cover 10, resulting in an improved removing means for an endoscope cover which simplifies the cleaning and sterilization of endoscopes.

25. It would have been obvious to one of ordinary skill in the art at the time of the invention to have provided the device of Silverstein with the structure of the tear thread 20 as taught by

Art Unit: 3739

Darras in order to have provided an improved removing means for the endoscope cover which simplifies the cleaning and sterilization of the endoscope to prevent the potential spread of viruses from one patient to another due to ineffective cleaning and sterilization of such instruments.

**Inserted image 1: Fig. 6 from Silverstein 4,646,722**



***Response to Amendment***

26. The amendment to claims 29, 41-43 and 45-48 in the response filed on March 2nd, 2010 is acknowledged.

27. Claims 29-48 remain pending in the application.

***Response to Arguments***

28. Applicant's arguments with respect to claims 29-48 have been considered but are moot in view of the new ground(s) of rejection. Regarding the Silverstein reference, the amendment to claims 29, 47, and 48 did not overcome the Silverstein reference. A new interpretation of the Silverstein reference has been laid out in the above rejections. With respect to the recitation,

Art Unit: 3739

“said single vacuum channel being a dedicated vacuum channel” the element 64 of Silverstein is a dedicated vacuum channel since gas is removed from element 64 (col. 7, lines 60-65). The limitations recited in the above discussed claimed have not been distinguished from the disclosure of the Silverstein reference.

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to ALIREZA NIA whose telephone number is (571)270-3076. The examiner can normally be reached on Mo.-Fri.-7:30 AM-5:00 PM EST-Alternate Fridays Off.

If attempts to reach the examiner by telephone are unsuccessful, the examiner’s supervisor, Linda C. Dvorak can be reached on 571-272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/A. N./  
Examiner, Art Unit 3739  
Alireza Nia  
April 6<sup>th</sup>, 2010

/Linda C Dvorak/  
Supervisory Patent Examiner, Art Unit  
3739

Application/Control Number: 10/541,297

Page 12

Art Unit: 3739